

Approved For Release 2005/02/17 : CIA-RDP78B04770A001400040006-8

MEMORANDUM FOR: *John R.*

Device appears to be simple and inexpensive rectifier, but it will probably require a lot of manual adjusting before any prints can be made. A photogrammetrist would have to set it up. I doubt if we would want to be bothered with all the work required in setting the machine up, but if a real requirement with high enough priority came through, this device could be effectively used (DATE)

Declass Review by NGA.

Approved For Release 2005/02/17 : CIA-RDP78B04770A001400040006-8